



EVPU

CE

CERTIFICATE OF CONFORMITY

No. 43040/101/1/2007/CE

Manufacturer

NEOMED, s.r.o.,

Gaštanová 2, 066 01 Humenné, Slovak Republic

Applicant

NEOMED, s.r.o.,

Gaštanová 2, 066 01 Humenné, Slovak Republic

Product

Dental unit

Types

**ARIA EXCELL S, ARIA EXCELL SE, ARIA EXCELL SR,
ARIA EXCELL TR, ARIA PRIMA S, ARIA PRIMA SE,
ARIA PRIMA SR, ARIA PRIMA TR**

The sample meets the essential safety requirements of the following directives:

Medical Devices Directive 93/42/EEC


A sample of the product was tested and found to be in conformity with following standards and / or technical specifications applied:

EN 60601-1: 1990 + A1; 1993 + A2; 1995 + A13: 1996	Test Report No. 0550B / 2007
EN ISO 7494-1: 2005	
EN ISO 6875: 1996	
EN 60601-1-2: 2001 + A1: 2006	Test Report No. 0550E / 2007

Certificate confirms that the equipment compiles with the Essentials requirements and objectives of mentioned directives / standards and refers only to the sample provided for testing / certification.

Nová Dubnica
October 16, 2007




Karol Glamoš
Director of NB No. 1293

EVPU a.s., NB No. 1293, Trenčianska 19, SK-018 51 Nová Dubnica, Slovak Republic
014059